


IV Pole Installation

1. Remove the antitipper or the tipping sleeve from the left side of the wheel chair.



2. Remove the nut and the bolt from the clamp on the IV Pole.

3. Slide the tube of the IV Pole over the tube of the wheelchair where the antitipper or the tipping sleeve was removed.



4. Install the nut and the bolt in the clamp on the IV Pole.



5. Install the antitipper or the tipping sleeve on the left side of the wheelchair.